

RUSH**RUSH****RUSH**

G & L Laboratories

246 Arlington St., Quincy, MA 02170
Tel: (617) 328-3663 Fax: (617) 472-0706

Lab. ID #

118608

To: Haverhill Board of Health
Attn: Richard McDonalds
4 Summer St., Room 210
Haverhill, MA 01830

Received Date/Time:	8/6/25, 2:50pm
Analyzed Date/Time:	8/6/25 3:00PM
Reported Date:	8/7/25
Collected By:	Client
Collection Date:	8/6/25

ANALYTICAL RESULTS

Sample#	Sampling Location	Beach Location ID#	Sampling Time	E. Coli (CFU/100mL)
1	Plugs Pond @ Main	3449	9:45AM ✓	36
2	Plugs Pond @ Side	5884	9:45AM ✓	48
Maximum Contaminant Level				235
Method of Reference				EPA 1603.1

Laboratory QC Results

QC	Unit	E. coli
Duplicate (D1)	CFU/100mL	65
Duplicate (D2)	CFU/100mL	65
Log Dif of Dup.		0.00
MS Recovery	%	93%
MS Rec Criteria	%	17%-117%
Dilution Blk.(Beginning)	CFU/100mL	Negative
Dilution Blk.(Ending)	CFU/100mL	Negative
Positive Control	CFU/100mL	Positive

Notes:

T.C.= 6.9 °C

Approved By:

Lab. Director

Date

8/7/25